



11302 Steeplecrest Dr. Houston, Texas 77065 Phone 832.320.7300 Fax 713.880.9886 www.ofite.com

PRESSURE TEST VERIFICATION

OFITE-Style Aging Cell, 500 mL, 300 Series SS

This is to certify that the OFITE product as serialized below has been hydraulically tested to the specified pressure at 75°F (23.9°C) for 5 minutes. This product is designed to operate at temperatures from 40°F (4.4°C) to 500°F (260°C) and a maximum allowable working pressure (MAWP) of 2,000 psi.

The pressure test for this product was conducted in accordance with OFITE's hydraulic test procedure ENG 105 with an NIST traceable gauge.

Serial Number:

Test Pressure:

(Tolerance Range: 3,475 to 3,525 PSI at 65° to 75° F)

Gauge Serial #:

Gauge Calibration Due:

Date:

Tested By:

Document Number:	OFI-ENG 105-300 Series SS	Pressure Test Certificate	Rev 1.0
Last Revised:	4/12/2024	Approved By:	ENG