



ES Meter Calibration Report

The Device Under Test is certified to meet the accuracy requirements stated in "API 13B-2, 11 Electrical Stability Test".

Order Number:	RP00000100
Serial # / Local I.D. #:	E004605
Calibration Date:	7/24/2026
*Recommended Due Date:	7/24/2027
Certificate Number:	E00460546227
Part Number:	131-50 ES Meter Calibration Report

Condition Found:	
Condition Left:	In Calibration
Repairs Required:	Yes
As Found Required:	No

OFI Testing Equipment, Inc. - 11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

CALIBRATION STANDARDS	Range: 609V - 1952V / Acceptable tolerance is $\pm 2.5\%$ of full scale (FS) at 25% - 75% FS Range, unless otherwise specified.					
Low Resistor Standard	Lot Number	Manufacturer	Description	Reading (V)	Tolerance	
	TF050R	Caddock*	Hi-Volt Resistor, 9.99 M Ω + / - 0.1 %	609 V	2 %	12.18 V
High Resistor Standard	Lot Number	Manufacturer	Description	Reading (V)	Tolerance	
	TF656R	Caddock*	Hi-Volt Resistor, 32.00 M Ω + / - 0.1 %	1952 V	2 %	39.04 V

These devices have been calibrated using instruments and known standards which are traceable to NIST.

AS LEFT MEASUREMENTS											
	Target	Cycle 1	Variance	Cycle 2	Variance	Cycle 3	Variance	Tolerance	Actual + / -	Difference	Pass / Fail
Low Resistor	609 V	618 V	9 V	618 V	9 V	618 V	9 V	± 12.18 V	9 V	9 V	Pass
High Resistor	1952 V	1924 V	-28 V	1924 V	-28 V	1924 V	-28 V	± 39.04 V	28 V	28 V	Pass
Probe in Water	≤ 3 V	1 V	-2 V	1 V	-2 V	1 V	-2 V	≤ 3 V	1 V	2 V	Pass
Probe in Air	∞	∞	∞	∞	∞	∞	∞	∞ Only	N/A	N/A	Pass
Probe Gap (Inches)	.061 in	.062 in	.001 in	.062 in	.001 in	.062 in	.001 in	.001 in	.001 in	.001 in	Pass

AS LEFT REPORT											
List all repairs made to the unit under AS LEFT measurements.											

Calibration Technician: Elmer Pumphrey

Technician Signature: *Elmer Pumphrey*

Date: 7/24/2026

Reviewed By: Samantha Korenek

Reviewer Signature: *Samantha Korenek*

Date: 7/28/2026