



Silverson Mixer Calibration Report

Order Number: RP00000003
Part Number: 152-16
Serial Number: 41180
I.D. Number:

Verification Date: 8/6/2026
Verification Due Date: 8/6/2027

Rotational Speed Test

Target Speed	Unit Display Speed	Tachometer Speed	Allowed Deviation	Actual Deviation	Pass / Fail
2500 RPM	2500 RPM	2516 RPM	± 300 RPM	16 RPM	Pass
5000 RPM	5000 RPM	5021 RPM	± 300 RPM	21 RPM	Pass
7500 RPM	7500 RPM	7524 RPM	± 300 RPM	24 RPM	Pass
9200 RPM	9200 RPM	9254 RPM	± 300 RPM	54 RPM	Pass

Calibrated Tachometer I.D. 680

Model Number: R7050
Serial Number: 220326689

Calibration Due Date: 7/17/2027

Calibrated By: Magdy Beshay

Date: 8/6/2026

Approved By: Samantha Korenek

Date: 8/6/2026

