



Order Number:	NEW
Serial # / Local I.D. #:	S2617300001
Calibration Date:	8/25/2026
Recommended Due Date:	8/25/2027
Certificate Number:	S261730000146259
Condition Found:	NEW
Condition Left:	In Calibration

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

171-49-1 Temperature Controller Calibration Report

Thermocouple Verification Check						
Channel	Target Temperature	As Found Temperature	Deviation Allowed	Adjusted Temperature	Deviation Allowed	Pass / Fail
Channel 1	120.0 °F	118.0 °F	± 4.0 °F	120.0 °F	± 0.0 °F	Pass
	450.0 °F	444.0 °F	± 10.0 °F	450.0 °F	± 0.0 °F	Pass
Channel 1 Thermocouple Serial Number or I.D. Number*				106DE3E		
Dry Block I.D. Number		674	Certificate Number		C-56620	
Model Number		ETC-400A 11	Calibration Date		10/27/2025	
Serial Number		692360-02363	Calibration Due Date		10/27/2026	
*If Applicable - If the Thermocouple is not serialized then enter N/A in the box .						

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

Shawn Maglitto
Calibration Technician

Technician Signature

8/25/2026
Date

Samantha Korenek
Reviewed By

Reviewer Signature

8/25/2026
Date