



Roller Oven Calibration Report

Order Number:	NEW
Serial # / Local I.D. #:	S2618040001
Calibration Date:	8/25/2026
Recommended Due Date:	8/25/2027
Certificate Number:	S261804000146259
Manufacturer:	OFI Testing Equipment
Part Number:	172-00-C

Condition Found:	NEW
Condition Left:	In Calibration
Repairs Required:	No
Static Oven:	No
(F) or (C):	(°F)
Max Temp:	500
As Found Required:	No

OFI Testing Equipment, Inc. - 11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

Calibration Equipment							
Thermometer	705	EA10	Z414473		C-67199	7/17/2026	7/17/2027
	I.D. Number	Model Number	Serial Number		Certificate Number	Calibration Date	Calibration Due Date
Tachometer	312	461920	12050313		C-54328	9/5/2025	9/5/2026
	I.D. Number	Model Number	Serial Number		Certificate Number	Calibration Date	Calibration Due Date
Ambient Monitor	169	EA20	Z375173		C-64913	5/20/2026	5/20/2027
	I.D. Number	Model Number	Serial Number		Certificate Number	Calibration Date	Calibration Due Date
Ambient Conditions	Lab / Room Temperature		75	(°F)	Lab / Room Humidity		53.0 %

AS LEFT MEASUREMENTS		(°F)	Max Oven Temp	500	
Set point	Process Temperature	Accepted Deviation	Actual Deviation	Pass/Fail	
100.0	99.0	5.0	-1.0	Pass	
	Pass				
200.0	203.0	5.0	3.0	Pass	
	Pass				
300.0	303.0	5.0	3.0	Pass	
	Pass				
400.0	401.0	5.0	1.0	Pass	
	Pass				
500.0	496.0	5.0	-4.0	Pass	
	Pass				
25 RPM	24.1	2	-0.9	Pass	
	Pass				
					Pass

Calibration Technician: Shawn Maglitta

Technician Signature: *Shawn Maglitta*

Date: 8/25/2026

Reviewed By: Samantha Korenek

Reviewer Signature: *Samantha Korenek*

Review Date: 8/25/2026

