



Service Provided To:

Ofite

Certificate Number:

S2617840002

Order Number:

1784

I.D. Number:

N/A

Serial Number:

S2617840002

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

Certificate of Calibration

Model	11B-200-OFI	Status	New
Manufacturer	Winters	Calibration Date	8/25/2026
Item Description	143-01-C	Calibration Due Date	8/25/2027
Pressure Range (psi)	200	Certificate Issued Date	8/26/2026
Accuracy (%FS)	2.00%	Calibration Interval	1 year
Temperature	73.4	Condition As Inspected	New
Humidity	55.5	Condition/ Data Results	Pass

Accuracy (%FS) - Tolerance Levels

Test Point / Nominal	Low (-) Tolerance	High (+) Tolerance	As found (DUT)	As Left (Maintained)	
0.00	-4.00	4.00	0.00	0.00	

PSI Actual Reading - Increasing (psi)

Test Point / Nominal	Low (-) Tolerance	High (+) Tolerance	As found (DUT)	As Left (Maintained)	Uncertainty	Condition
0.00	0.00	0.00	0.00	0.00	±34 psi	Pass
50.0	46.00	54.00	53.00	53.00	±34 psi	Pass
100.0	96.00	104.00	102.00	102.00	±34 psi	Pass
150.0	146.00	154.00	152.00	152.00	±34 psi	Pass
200.00	196.00	204.00	203.00	203.00	±34 psi	Pass

PSI Actual Reading - Decreasing (psi)

Test Point / Nominal	Low (-) Tolerance	High (+) Tolerance	As found (DUT)	As Left (Maintained)	Uncertainty	Condition
150.0	146.00	154.00	152.00	152.00	±34 psi	Pass
100.0	96.00	104.00	102.00	102.00	±34 psi	Pass
50.0	46.00	54.00	52.00	52.00	±34 psi	Pass
0.00	0.00	0.00	0.00	0.00	±34 psi	Pass

Standards Used			
ID# / Model Used	Description	Cal Date	Due Date
ADT 673-05-GP15K-PSI-N	DIGITAL PRESSURE CALIBRATOR	4/1/2026	3/31/2027

Internal Calibration Procedure
SOP 7.2 Calibration Methods Rev 0 Dated: 4/30/26

Calibrations within the certificate are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Intervals of calibration for recall are determined by the end user. Calibration results relate only to items calibrated and apply to the items as received. The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%. The measurement uncertainty has been considered using simple acceptance-where acceptance limit is the same as the tolerance limit. The Test Uncertainty Ratio (TUR) shall be greater than 4:1 whenever possible, and no less than 1:1. This is a shared risk decision rule in which the customer also has responsibility for determining acceptance of the result.

		
<u>Benjamin McMullin</u>	_____	<u>8/26/2026</u>
Calibration Technician	Technician Signature	Date
		
<u>Samantha Korenek</u>	_____	<u>8/26/2026</u>
Reviewed By	Reviewer Signature	Date