



Order Number:	New
Serial # / Local I.D. #:	S2616380001
Calibration Date:	8/19/2026
Recommended Due Date:	8/19/2027
Certificate Number:	S261638000146253
Condition Found:	NEW
Condition Left:	In Calibration

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

130-83-1 Model 3025 Viscometer 220V Calibration Report

Thermocouple Calibration Check			°F or °C			°F
Target Temperature	100	300	500	Allowed Deviation	Actual Deviation	Pass / Fail
Amulet Reading	100.2	300.1	499.9	2	0.2	Pass
	Pass	Pass	Pass			
Software Reading	100	300.2	500.1	2	0.2	Pass
	Pass	Pass	Pass			
Dry Block I.D. Number	44		Certificate Number		C-51136	
Model Number	ETC400A		Calibration Date		6/15/2026	
Serial Number	623915-01434		Calibration Due Date		6/15/2027	

Pressure Gauge Calibration Check						
Target Pressure	Master Gauge Reading	Software Reading	Unit Gauge Reading	Allowed Deviation	Actual Deviation	Pass / Fail
650	655 PSI	653 PSI	650 PSI	± 50 PSI	5 PSI	Pass
		Pass	Pass			
1250	1248 PSI	1256 PSI	1250 PSI	± 50 PSI	6 PSI	Pass
		Pass	Pass			
1900	1909 PSI	1912 PSI	1900 PSI	± 50 PSI	12 PSI	Pass
		Pass	Pass			
Master Gauge I.D. Number		334	Certificate Number		C-58777	
Model Number	M1 series		Calibration Date		1/6/2026	
Serial Number	703541		Calibration Due Date		1/6/2027	

ORCADA Calibration Information			
R^2 Result Value		R^2 Allowed Deviation	
1.0000		Low	High
		0.9990	1.0000
Pass			
Calibration Fluid Information			
200 cP Lot Number	Activation Date		Expiration Date
510501	8/5/2026		9/30/2028

Motor Speed Accuracy Check				
Target Speed	Tachometer Measured Reading	Allowed Deviation	Actual Deviation	Pass / Fail
600.0 RPM	599.9 RPM	± 6.0 RPM	0.1 RPM	Pass
	Pass			
300.0 RPM	300.1 RPM	± 3.0 RPM	0.1 RPM	Pass
	Pass			
200.0 RPM	199.9 RPM	± 2.0 RPM	0.1 RPM	Pass
	Pass			
100.0 RPM	100.1 RPM	± 1.0 RPM	0.1 RPM	Pass
	Pass			
6.0 RPM	06.1 RPM	± 1.0 RPM	0.1 RPM	Pass
	Pass			
3.0 RPM	03.0 RPM	± 1.0 RPM	0.0 RPM	Pass
	Pass			
Tachometer I.D. Number		706	Certificate Number	C-54329
Model Number		R7050	Calibration Date	9/5/2025
Serial Number		220502483	Calibration Due Date	9/5/2026

This report meets or exceeds the requirements outlined in API RP 10B-2 Recommended Practice for Testing Well Cements for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

Elvin Le
Calibration Technician



Technician Signature

8/19/2026

Date

Therese Ikladious
Reviewed By



Reviewer Signature

8/19/2026

Date