



Chiller Calibration Report

Order Number: SV#00000061
Serial Number: 0152657801161210
Model Number: RD-DA-123/124

Calibration Date: 8/19/2026
Calibration Due Date: 8/19/2027
Manufacturer: THERMO SCI

Thermocouple Calibration Check:			°F / °C	°C	
Target Temperature	Unit Reading	Master Thermometer	Accepted Deviation	Actual Deviation	Pass / Fail
113	113	112.8	2.2	0.2	Pass
35	35	35.3	2.2	0.3	Pass
25	24.6	25.1	2.2	0.5	Pass

Master Thermocouple Information: 668

Model Number: 90080-09
Serial Number: 181334600

Calibration Due Date: 1/6/2027

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: Jeffrey Vo

Date: 8/19/2026

Approved By: Samantha Korenek

Date: 8/27/2026