



Pressure Station Accuracy Report

Work Order: SV#00000061

Calibration Date: 8/18/2026

Pressure Station: RD-DA-001, 002

Calibration Due: 8/18/2027

Number of Master Gauges Used: 1

Master Gauge 1		Range:	300
I.D. Number	Serial Number	Model Number	Calibration Due
130	162695	M1 series	9/5/2026

Number of Gauges to be Tested: 1

Test Gauge 1		Max Pressure on Unit Gauge (PSI)		200
Target Pressure	Master Gauge Reading	Unit Gauge Reading	Accepted Deviation	Pass / Fail
50	51 PSI	50 PSI	5 PSI	Pass
100	99 PSI	100 PSI	5 PSI	Pass
150	146 PSI	150 PSI	5 PSI	Pass
Gauge Location		Left	Master Gauge Used	130

The gauges tested above were tested for compliance with API 13A, 5.2.7 "Laboratory pressure-measuring device". When passed, these gauges are verified to measure within 2 ½% of Full Scale when compared to an NIST traceable gauge, measured at 25%, 50% and 75% of Full Range.

Technician Tested: Jeffrey Vo

Date: 8/18/2026

Approved By: Samantha Korenek

Date: 8/27/2026