



Mixer Calibration Report

Order Number:	SV#00000061	Calibration Date:	8/18/2026
Part Number:	152-50	Calibration Due Date:	8/18/2027
Serial Number:	N/A	Manufacturer:	OFITE
I.D. Number:	RD-DA-039	Number of Spindles:	1
		Number of Speed Checks:	1

Rotational Speed Test				Max Speed of Unit		11500
Target Speed		Unit Display Speed	Tachometer Speed	Accepted Deviation	Actual Deviation	Pass / Fail
1	11500	11500	11654	288	154	Pass

Tachometer Calibration I.D. # 706

Model Number:	R7050	Calibration Due Date:	9/5/2026
Serial Number:	220502483		

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Calibrated By: Jeffrey Vo Date: 8/18/2026

Approved By: Samantha Korenek Date: 8/27/2026

