



IKA Mixer Calibration Report

Order Number: SV#00000061
Part Number: T 18 D S1
Serial Number: 03498808
I.D. Number: RD-DA-318

Calibration Date: 8/19/2026
Calibration Due Date: 8/19/2027
Manufacturer: IKA
Number of Spindles: 1
Number of Speed Checks: 5

Rotational Speed Test			Max Speed of Unit		10000
Target Speed	Unit Display Speed	Tachometer Speed	Accepted Deviation	Actual Deviation	Pass / Fail
2000	2000	1998	250	2	Pass
4000	4000	4140	250	140	Pass
6000	6000	6142	250	142	Pass
8000	8000	8004	250	4	Pass
10000	10000	10242	250	242	Pass

Tachometer Calibration I.D. # 706

Model Number: R7050
Serial Number: 220502483

Calibration Due Date: 9/5/2026

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Calibrated By: Jeffrey

Date: 8/19/2026

Approved By: Samantha Korenek

Date: 8/27/2026

