



Static Oven Calibration Report

Order Number:	SO#00000061
Serial # / Local I.D. #:	11-05157
Calibration Date:	8/18/2026
Recommended Due Date:	8/18/2027
Certificate Number:	11-0515746252
Manufacturer:	BINDER
ID Number:	RD-DA-105

Condition Found:	Out of Calibration
Condition Left:	In Calibration
Repairs Required:	No
Static Oven:	Yes
(F) or (C):	(°C)
Max Temp:	204
As Found Required:	No

OFI Testing Equipment, Inc. - 11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

Calibration Equipment						
Thermometer	154	EA10	Z375331	C-67201	7/17/2026	7/17/2027
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Ambient Monitor	261	EA20	Z375174	C-59375	1/20/2026	1/20/2027
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Ambient Conditions	Lab / Room Temperature		72.9	(°C)	Lab / Room Humidity	61.1 %

AS LEFT MEASUREMENTS		(°C)	Max Oven Temp	204	
Set point	Process Temperature	Accepted Deviation	Actual Deviation	Pass/Fail	
37.8	38.0	2.2	0.2	Pass	
	Pass				
93.3	95.0	2.2	1.7	Pass	
	Pass				
148.9	149.0	2.2	0.1	Pass	
	Pass				
					Pass

Calibration Technician: Jeffrey Vo

Date: 8/18/2026

Reviewed By: Samantha Korenek

Review Date: 8/27/2026

