



## Static Oven Calibration Report

|                          |             |
|--------------------------|-------------|
| Order Number:            | SO#00000061 |
| Serial # / Local I.D. #: | 119513      |
| Calibration Date:        | 8/19/2026   |
| Recommended Due Date:    | 8/19/2027   |
| Certificate Number:      | 11951346253 |
| Manufacturer:            | BLUE M      |
| ID Number:               | RD-DA-104   |

|                    |                    |
|--------------------|--------------------|
| Condition Found:   | Out of Calibration |
| Condition Left:    | In Calibration     |
| Repairs Required:  | No                 |
| Static Oven:       | Yes                |
| ( F ) or ( C ):    | (°C)               |
| Max Temp:          | 260                |
| As Found Required: | No                 |

OFI Testing Equipment, Inc. - 11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: [www.ofite.com](http://www.ofite.com)

| Calibration Equipment |                        |              |               |                    |                     |                      |
|-----------------------|------------------------|--------------|---------------|--------------------|---------------------|----------------------|
| Thermometer           | 154                    | EA10         | Z375331       | C-67201            | 7/17/2026           | 7/17/2027            |
|                       | I.D. Number            | Model Number | Serial Number | Certificate Number | Calibration Date    | Calibration Due Date |
| Ambient Monitor       | 261                    | EA20         | Z375174       | C-59375            | 1/20/2026           | 1/20/2027            |
|                       | I.D. Number            | Model Number | Serial Number | Certificate Number | Calibration Date    | Calibration Due Date |
| Ambient Conditions    | Lab / Room Temperature |              | 73.1          | (°C)               | Lab / Room Humidity | 61.5 %               |

| AS LEFT MEASUREMENTS |                     | (°C)               | Max Oven Temp    | 260       |  |
|----------------------|---------------------|--------------------|------------------|-----------|--|
| Set point            | Process Temperature | Accepted Deviation | Actual Deviation | Pass/Fail |  |
| 37.8                 | 37.0                | 2.2                | -0.8             | Pass      |  |
|                      | Pass                |                    |                  |           |  |
| 93.3                 | 92.0                | 2.2                | -1.3             | Pass      |  |
|                      | Pass                |                    |                  |           |  |
| 148.9                | 151.0               | 2.2                | 2.1              | Pass      |  |
|                      | Pass                |                    |                  |           |  |
| 204.4                | 203.0               | 2.2                | -1.4             | Pass      |  |
|                      | Pass                |                    |                  |           |  |
| 260.0                | 261.0               | 2.2                | 1.0              | Pass      |  |
|                      | Pass                |                    |                  |           |  |

Calibration Technician: Jeffrey Vo

Date: 8/19/2026

Reviewed By: Samantha Korenek

Review Date: 8/27/2026

