



Dependable Products From People You Trust

11302 Steeplecrest Drive, Houston, Texas 77065 Phone (713)880-9885 Fax (713)880-9886 Web www.ofite.com

OFI TESTING EQUIPMENT, INC.

CERTIFICATE OF CONFORMANCE

This is to certify that the following instrument has been built and tested by OFI Testing Equipment, Inc., and will conform to the specifications as outlined in API Recommended Practice Publication API RP 13B-1.

PART NO.	<u>170-00-4</u>
DESCRIPTION	<u>HTHP FILTER PRESS , 4 UNIT, 115 V</u>
UNIT SERIAL NUMBER	<u></u>
TOLERANCE	<u>375°F ± 25 (350 - 400° F)</u>
LP PRV SET PRESSURE	<u></u>
HP PRV SET PRESSURE	<u></u>
CELL 1 SERIAL NUMBER	<u></u>
CELL 2 SERIAL NUMBER	<u></u>
CELL 3 SERIAL NUMBER	<u></u>
CELL 4 SERIAL NUMBER	<u></u>
ASSEMBLED BY	<u></u>
TESTED BY	<u></u>
DATE TESTED	<u></u>

THIS HIGH PRESSURE SIDE OF THIS UNIT WAS PRESSURE TESTED TO 1350 PSI THE LOW PRESSURE SIDE OF THIS UNIT WAS PRESURE TESTED TO 750 PSI AND INSPECTED FOR LEAKAGE. HEATING JACKETS WERE TESTED TO THE HEAT RANGE OF 385 - 465° F.