



Static Oven Calibration Report

Order Number:	SO#00000061
Serial # / Local I.D. #:	13-21321
Calibration Date:	8/18/2026
Recommended Due Date:	8/18/2027
Certificate Number:	13-2132146252
Manufacturer:	BINDER
ID Number:	RD-DA-102

Condition Found:	Out of Calibration
Condition Left:	In Calibration
Repairs Required:	No
Static Oven:	Yes
(F) or (C):	(°C)
Max Temp:	260
As Found Required:	No

OFI Testing Equipment, Inc. - 11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

Calibration Equipment						
Thermometer	154	EA10	Z375331	C-67201	7/17/2026	7/17/2027
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Ambient Monitor	261	EA20	Z375174	C-59375	1/20/2026	1/20/2027
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Ambient Conditions	Lab / Room Temperature		72.7	(°C)	Lab / Room Humidity	61.3 %

AS LEFT MEASUREMENTS		(°C)	Max Oven Temp	260	
Set point	Process Temperature	Accepted Deviation	Actual Deviation	Pass/Fail	
37.8	38.0	2.2	0.2	Pass	
	Pass				
93.3	92.0	2.2	-1.3	Pass	
	Pass				
148.9	148.0	2.2	-0.9	Pass	
	Pass				
204.4	206.0	2.2	1.6	Pass	
	Pass				
260.0	259.0	2.2	-1.0	Pass	
	Pass				

Calibration Technician: Jeffrey Vo

Date: 8/18/2026

Reviewed By: Samantha Korenek

Review Date: 8/27/2026

