



Static Oven Calibration Report

Order Number:	SO#00000061
Serial # / Local I.D. #:	14-01273
Calibration Date:	8/18/2026
Recommended Due Date:	8/18/2027
Certificate Number:	14-0127346252
Manufacturer:	BINDER
ID Number:	RD-DA-101

Condition Found:	Out of Calibration
Condition Left:	In Calibration
Repairs Required:	No
Static Oven:	Yes
(F) or (C):	(°C)
Max Temp:	260
As Found Required:	No

OFI Testing Equipment, Inc. - 11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

Calibration Equipment						
Thermometer	154	EA10	Z375331	C-67201	7/17/2026	7/17/2027
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Ambient Monitor	261	EA20	Z375174	C-59375	1/20/2026	1/20/2027
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Ambient Conditions	Lab / Room Temperature		72.9	(°C)	Lab / Room Humidity	61.1 %

AS LEFT MEASUREMENTS		(°C)	Max Oven Temp	260	
Set point	Process Temperature	Accepted Deviation	Actual Deviation	Pass/Fail	
37.8	38.0 Pass	2.2	0.2	Pass	
93.3	94.0 Pass	2.2	0.7	Pass	
148.9	149.0 Pass	2.2	0.1	Pass	
204.4	206.0 Pass	2.2	1.6	Pass	
260.0	261.0 Pass	2.2	1.0	Pass	
	Pass				

Calibration Technician: Jeffrey Vo

Date: 8/18/2026

Reviewed By: Samantha Korenek

Review Date: 8/27/2026

