



Static Oven Calibration Report

Order Number:	SO#00000061
Serial # / Local I.D. #:	1585080190101
Calibration Date:	8/19/2026
Recommended Due Date:	8/19/2027
Certificate Number:	158508019010146253
Manufacturer:	COLE-PARMER
ID Number:	RD-DA-106

Condition Found:	Out of Calibration
Condition Left:	In Calibration
Repairs Required:	No
Static Oven:	Yes
(F) or (C):	(°C)
Max Temp:	260
As Found Required:	No

OFI Testing Equipment, Inc. - 11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

Calibration Equipment						
Thermometer	154	EA10	Z375331	C-67201	7/17/2026	7/17/2027
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Ambient Monitor	261	EA20	Z375174	C-59375	1/20/2026	1/20/2027
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Ambient Conditions	Lab / Room Temperature		67.5	(°C)	Lab / Room Humidity	64.1 %

AS LEFT MEASUREMENTS		(°C)	Max Oven Temp	260	
Set point	Process Temperature	Accepted Deviation	Actual Deviation	Pass/Fail	
37.8	39.0 Pass	2.2	1.2	Pass	
93.3	95.0 Pass	2.2	1.7	Pass	
148.9	149.0 Pass	2.2	0.1	Pass	
204.4	206.0 Pass	2.2	1.6	Pass	
260.0	259.0 Pass	2.2	-1.0	Pass	
				Pass	

Calibration Technician: Jeffrey Vo

Date: 8/19/2026

Reviewed By: Samantha Korenek

Review Date: 8/27/2026

