



Silverson Mixer Calibration Report

Order Number: SV#00000061
Part Number: L4R
Serial Number: 18361
I.D. Number: RD-DA-011

Verification Date: 8/18/2026
Verification Due Date: 8/18/2027

Rotational Speed Test

Target Speed	Unit Display Speed	Tachometer Speed	Allowed Deviation	Actual Deviation	Pass / Fail
2500 RPM	2500 RPM	2422 RPM	± 300 RPM	78 RPM	Pass
5000 RPM	5000 RPM	4946 RPM	± 300 RPM	54 RPM	Pass
7500 RPM	7500 RPM	7562 RPM	± 300 RPM	62 RPM	Pass
9200 RPM	8900 RPM	8998 RPM	± 300 RPM	300 RPM	Pass

Calibrated Tachometer I.D. 706

Model Number: R7050
Serial Number: 220502483

Calibration Due Date: 9/5/2026

Calibrated By: Jeffrey Vo

Date: 8/18/2026

Approved By: Samantha Korenek

Date: 8/27/2026

