



Chiller Calibration Report

Order Number: SV#00000061
Serial Number: 3G12A0396
Model Number: RD-DA-122

Calibration Date: 8/19/2026
Calibration Due Date: 8/19/2027
Manufacturer: POLYSCI

Thermocouple Calibration Check:			°F / °C	°C	
Target Temperature	Unit Reading	Master Thermometer	Accepted Deviation	Actual Deviation	Pass / Fail
30	30.4	31.1	2.2	1.1	Pass
20	20.3	20.1	2.2	0.3	Pass
10	9.9	10.2	2.2	0.3	Pass

Master Thermocouple Information: 668

Model Number: 90080-09
Serial Number: 181334600

Calibration Due Date: 1/6/2027

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: Jeffrey Vo

Date: 8/19/2026

Approved By: Samantha Korenek

Date: 8/27/2026