



120-60 Constant Speed Blender 115v Calibration Report

Sales/ Work Order: SV#00000061 Calibration Date: 8/18/2026
Serial Number: 5145060000001 Recommended Calibration Date: 8/18/2027
Blender SN: 5145060000001
ID Number: RD-DA-056

Rotational Speed Test:					
Target Speed	Low Point Actual Speed	High Point Actual Speed	Accepted Deviation	Actual Deviation	Pass / Fail
4000 RPM	3997 RPM	4086 RPM	± 100 RPM	89 RPM	Pass
12000 RPM	11996 RPM	12045 RPM	± 250 RPM	49 RPM	Pass

Tachometer Calibration Information: 706

Model Number: R7050 Calibration Due Date: 9/5/2026
Serial Number: 220502483

This report meets or exceeds the requirements outlined in API RP 10B-2 Recommended Practice for Testing Well Cements for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: Jeffrey Vo Date: 8/18/2026

Approved By: Samantha Korenek Date: 8/27/2026