



Caframo Mixer Calibration Report

Order Number: SV#00000061
Part Number: 6015
Serial Number: 6001660
I.D. Number: RD-DA-318

Calibration Date: 8/19/2026
Calibration Due Date: 8/19/2027
Manufacturer: IKA
Number of Spindles: 1
Number of Speed Checks: 3

Rotational Speed Test			Max Speed of Unit		10000
Target Speed	Unit Display Speed	Tachometer Speed	Accepted Deviation	Actual Deviation	Pass / Fail
2500	2500	2531	250	31	Pass
5000	5000	5089	250	89	Pass
7500	7500	7468	250	32	Pass

Tachometer Calibration I.D. # 706

Model Number: R7050
Serial Number: 220502483

Calibration Due Date: 9/5/2026

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Calibrated By: Jeffrey

Date: 8/19/2026

Approved By: Samantha Korenek

Date: 8/27/2026

