



Pressure Station Accuracy Report

Work Order: SV#00000061

Calibration Date: 8/19/2026

Pressure Station: RD-DA-086

Calibration Due: 8/19/2027

Number of Master Gauges Used: 1

Master Gauge 1		Range:	3000
I.D. Number	Serial Number	Model Number	Calibration Due
334	703541	M1 series	1/6/2027

Number of Gauges to be Tested: 1

Test Gauge 1		Max Pressure on Unit Gauge (PSI)		1100
Target Pressure	Master Gauge Reading	Unit Gauge Reading	Accepted Deviation	Pass / Fail
275	268 PSI	280 PSI	28 PSI	Pass
550	560 PSI	560 PSI	28 PSI	Pass
825	825 PSI	840 PSI	28 PSI	Pass
Gauge Location		Master Gauge Used		

The gauges tested above were tested for compliance with API 13A, 5.2.7 "Laboratory pressure-measuring device". When passed, these gauges are verified to measure within 2 ½% of Full Scale when compared to an NIST traceable gauge, measured at 25%, 50% and 75% of Full Range.

Technician Tested: Jeffrey Vo

Date: 8/19/2026

Approved By: Samantha Korenek

Date: 8/27/2026