



## Roller Oven Calibration Report

|                          |                       |
|--------------------------|-----------------------|
| Order Number:            | SV00000061            |
| Serial # / Local I.D. #: | 13-022                |
| Calibration Date:        | 8/19/2026             |
| Recommended Due Date:    | 8/19/2027             |
| Certificate Number:      | 13-02246253           |
| Manufacturer:            | OFI Testing Equipment |
| Part Number:             | 173-00-RC             |

|                    |                    |
|--------------------|--------------------|
| Condition Found:   | Out of Calibration |
| Condition Left:    | In Calibration     |
| Repairs Required:  | No                 |
| Static Oven:       | No                 |
| ( F ) or ( C ):    | (°F)               |
| Max Temp:          | 400                |
| As Found Required: | No                 |

OFI Testing Equipment, Inc. - 11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

| Calibration Equipment |                        |              |               |                    |                     |                      |
|-----------------------|------------------------|--------------|---------------|--------------------|---------------------|----------------------|
| Thermometer           | 257                    | EA10         | Z375329       | C-60756            | 2/25/2026           | 2/25/2027            |
|                       | I.D. Number            | Model Number | Serial Number | Certificate Number | Calibration Date    | Calibration Due Date |
| Tachometer            | 679                    | R7050        | 220326766     | C-67204            | 7/17/2026           | 7/17/2027            |
|                       | I.D. Number            | Model Number | Serial Number | Certificate Number | Calibration Date    | Calibration Due Date |
| Ambient Monitor       | 261                    | EA20         | Z375174       | C-59375            | 1/20/2026           | 1/20/2027            |
|                       | I.D. Number            | Model Number | Serial Number | Certificate Number | Calibration Date    | Calibration Due Date |
| Ambient Conditions    | Lab / Room Temperature |              | 71.6          | (°F)               | Lab / Room Humidity | 50.1 %               |

| AS LEFT MEASUREMENTS |                     | (°F)               | Max Oven Temp    | 400 |           |
|----------------------|---------------------|--------------------|------------------|-----|-----------|
| Set point            | Process Temperature | Accepted Deviation | Actual Deviation |     | Pass/Fail |
| 100.0                | 103.0               | 5.0                | 3.0              |     | Pass      |
|                      | Pass                |                    |                  |     |           |
| 200.0                | 204.0               | 5.0                | 4.0              |     | Pass      |
|                      | Pass                |                    |                  |     |           |
| 300.0                | 304.0               | 5.0                | 4.0              |     | Pass      |
|                      | Pass                |                    |                  |     |           |
| 400.0                | 404.0               | 5.0                | 4.0              |     | Pass      |
|                      | Pass                |                    |                  |     |           |
| 25 RPM               | 25.9                | 2                  | 0.9              |     | Pass      |
|                      | Pass                |                    |                  |     |           |
|                      |                     |                    |                  |     | Pass      |

|                         |                  |
|-------------------------|------------------|
| Calibration Technician: | Magdy Beshay     |
| Date:                   | 8/19/2026        |
| Reviewed By:            | Samantha Korenek |
| Review Date:            | 8/27/2026        |

