



EP and Lubricity Tester Calibration Report

Order Number: SV00000064
Serial Number: 5089010000001
Part Number: N/A
Item I.D. #:

Calibration Date: 8/19/2026
Recommended Calibration Date: 8/19/2027
Manufacturer: Ofite

RPM Calibration Check				
Target Speed	Tachometer Reading (RPM)	Allowed Deviation	Actual Deviation	Pass / Fail
60 RPM	60.0	± 1.5 RPM	.0 RPM	Pass
200 RPM	200.0	± 5.0 RPM	.0 RPM	Pass
600 RPM	599.0	± 15.0 RPM	1.0 RPM	Pass
1000 RPM	994.0	± 25.0 RPM	6.0 RPM	Pass

Tachometer Calibration Information:

679

Model Number: R7050
Serial Number: 220326766

Calibration Due Date: 7/17/2027

Deionized Water Check				
Target Torque (in-lb)	Unit Reading (in-lb)	Acceptable Deviation (in-lb)	Actual Deviation	Pass / Fail
34.0 in-lb	35.2	± 2.0 in -lb	1.2 in-lb	Pass

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: Magdy Beshay

Date: 8/19/2026

Approved By: Samantha Korenek

Date: 8/27/2026