



Order Number:	NEW
Serial # / Local I.D. #:	S2618920001
Calibration Date:	9/2/2026
Recommended Due Date:	9/2/2027
Certificate Number:	S261892000146267
Condition Found:	NEW
Condition Left:	In Calibration
Blender S/N:	2600091

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

120-64-1-F Constant Speed Blender 230v Calibration Report

Rotational Speed Test					
Target Speed	Low Point Actual Speed	High Point Actual Speed	Accepted Deviation	Actual Deviation	Pass / Fail
500 RPM	498 RPM	502 RPM	± 10 RPM	4 RPM	Pass
	Pass	Pass			
2000 RPM	1997 RPM	2002 RPM	± 40 RPM	5 RPM	Pass
	Pass	Pass			
Tachometer I.D. Number		173	Certificate Number	C-59863	
Model Number		461920	Calibration Date	2/3/2026	
Serial Number		131105905	Calibration Due Date	2/3/2027	

This report meets or exceeds the requirements outlined in API RP 10B-2 Recommended Practice for Testing Well Cements for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

Ebram Azer
Calibration Technician

Ebram Azer
Technician Signature

9/2/2026
Date

Samantha Korenek
Reviewed By

Samantha Korenek
Reviewer Signature

9/2/2026
Date