



OFI Testing Equipment, Inc.
 11302 Steeplecrest Drive Houston, TX 77065 USA
 Phone: 832-320-7300 · Fax: 713-880-9886 · Web: www.ofite.com

Roller Oven Series Calibration Report

Repairs Required? No

As Found Data Required by Customer? No

Static Oven: No

(°F) or (°C) : (°F)

Part Number: 173-00-RC

Manufacturer: OFI Testing Equipment

Serial Number: 13-01

S/O# or I/O# SV00000068

Master Testing Equipment							
Thermometer:	<u>257</u>	Model #:	<u>EA10</u>	Serial #:	<u>Z375329</u>	Calibration Due:	<u>2/25/2027</u>
Tachometer:	<u>679</u>	Model #:	<u>R7050</u>	Serial #:	<u>220326766</u>	Calibration Due:	<u>7/17/2027</u>

Ambient Conditions			
Ambient Monitor:	<u>261</u>	Model No:	<u>EA20</u>
Calibration Date:	<u>1/20/2026</u>	Calibration Due:	<u>1/20/2027</u>
Serial No:	<u>Z375174</u>		
Ambient Conditions:	Room Temperature:	<u>70</u>	Relative Humidity: <u>49</u>

AS LEFT MEASUREMENTS		(°F)		
Set point	Process Temperature	Accepted Deviation	Actual Deviation	Pass/Fail
100	100	5	0	Pass
200	200	5	0	Pass
300	297	5	-3	Pass
400	399	5	-1	Pass
25 RPM	25.3	2	0.3	Pass
				Pass

Technician: Arnulfo Muniz

Calibration Date: 9/1/2026

Recommended Calibration Date *: 9/1/2027
 * To be provided by customer

Reviewed By: Antonio Rico

Review Date: 9/2/2026

