



OFI Testing Equipment, Inc.
 11302 Steeplecrest Drive Houston, TX 77065 USA
 Phone: 832-320-7300 · Fax: 713-880-9886 · Web: www.ofite.com

Roller Oven Series Calibration Report

Repairs Required? No

As Found Data Required by Customer? No

Static Oven: Yes

(°F) or (°C) : (°F)

Part Number: 173-00-C

Manufacturer: Delux

I.D. Number:

Serial Number: 1412201

S/O# or I/O# SV00000068

Master Testing Equipment

Thermometer: 257 Model #: EA10 Serial #: Z375329 Calibration Due: 2/25/2027

Ambient Conditions

Ambient Monitor: 261 Model No: EA20

Calibration Date: 1/20/2026

Calibration Due: 1/20/2027

Serial No: Z375174

Ambient Conditions: Room Temperature: 70

Relative Humidity: 49

AS LEFT MEASUREMENTS		(°F)		
Set point	Process Temperature	Accepted Deviation	Actual Deviation	Pass/Fail
100	100	5	0	Pass
200	198	5	-2	Pass
300	296	5	-4	Pass
400	396	5	-4	Pass
				Pass

Technician: Arnulfo Muniz

Calibration Date: 9/1/2026

Recommended Calibration Date *: 9/1/2027

* To be provided by customer

Reviewed By: Antonio Rico

Review Date: 9/2/2026

