



Pressure Gauge Verification Report

Work Order: SV00000068

Verification Date: 9/1/2026

Equipment: 4

Verification Due: 9/1/2027

Number of Master Gauges Used: 1

Master Gauge 1		Range:	3000
I.D. Number	Serial Number	Model Number	Calibration Due
126	5186403001	68935-26	5/23/2027

Number of Gauges to be Tested: 1

Test Gauge 1		Max Pressure on Unit Gauge (PSI)		200
Target Pressure	Master Gauge Reading	Unit Gauge Reading	Accepted Deviation	Pass / Fail
50	51 PSI	50 PSI	2 PSI	Pass
100	98 PSI	100 PSI	2 PSI	Pass
150	151 PSI	150 PSI	2 PSI	Pass
Gauge Location		RIGHT	Master Gauge Used	126

The gauges tested above were tested for compliance with API 13A, 5.2.7 "Laboratory pressure-measuring device". When passed, these gauges are verified to measure within 2 ½% of Full Scale when compared to an NIST traceable gauge, measured at 25%, 50% and 75% of Full Range.

Technician Tested: Arnolfo Muniz

Date: 9/1/2026

Approved By: Antonio Rico

Date: 9/2/2026