



Order Number:	RP00000169
Serial # / Local I.D. #:	5188740000002
Verification Date:	9/2/2026
Recommended Due Date:	9/2/2027
Certificate Number:	
Condition Found:	Out of Tolerance
Condition Left:	In Tolerance
As Found Required:	No

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

Pressure Gauge Verification Report

As Left		Gauge Location / Assembly Part Number		170-06	
Max Pressure on Unit Gauge (PSI)				200 PSI	
Target Pressure	Master Gauge Reading	Unit Gauge Reading	Allowed Deviation	Actual Deviation	Pass / Fail
50 PSI	49 PSI	50 PSI	± 2 PSI	1 PSI	Pass
100 PSI	99 PSI	100 PSI	± 2 PSI	1 PSI	Pass
150 PSI	151 PSI	150 PSI	± 2 PSI	-1 PSI	Pass
Master Guage I.D. Number		166	Certificate Number		C-58778
Model Number		68935-22	Calibration Date		1/6/2026
Serial Number		5186402001	Calibration Due Date		1/6/2027

The gauge tested above was tested for compliance with API 13A, 5.2.7 "Laboratory pressure-measuring device". When passed, this gauge is verified to measure within 2 ½% of Full Scale when compared to an NIST traceable gauge, measured at 25%, 50% and 75% of Full Range.

Magdy Beshay
Verification Technician

Samantha Korenek
Reviewed By

9/2/2026
Date

9/8/2026
Date