



Order Number:	New
Serial # / Local I.D. #:	S2616120002
Calibration Date:	9/9/2026
Recommended Due Date:	9/9/2027
Certificate Number:	S261612000246274
Condition Found:	NEW
Condition Left:	In Calibration

11302 Steeplecrest Dr - Houston, Texas - 77064 - Phone: 832.320.7300 - E-Mail: www.ofite.com

294-50 Capillary Suction Timer Calibration Report

Balance Calibration Check **							
Target Weight	Reading (Seconds)	Temperature (F°)		Allowed Deviation	Actual Deviation	Pass / Fail	
Reading 1	10.9 Sec	74.5 F°		± 3.6 Sec	0.1 Sec	Pass	
	Pass	Pass					
Reading 2	11.0 Sec	74.5 F°		± 3.6 Sec			
	Pass	Pass					
Reading 3	11.0 Sec	74.5 F°		± 3.6 Sec			
	Pass	Pass					
Calibration Fluid Information *		Calibration Fluid Lot #		Calibration Fluid Exp.Date			
10000 µS		6061925		6/30/2027			
Thermometer I.D.		668	Certificate Number		C-58767		
Model Number		90080-09		Calibration Date		1/6/2026	
Serial Number		181334600		Calibration Due Date		1/6/2027	

* Total dissolved of solids = 6600 Parts Per Million (PPM)

** Three tests of the same temperature should produce a deviation of no more than ± 3.6 seconds.

Emad Hanna
Calibration Technician

Emad Hanna
Technician Signature

9/9/2026
Date

Samantha Korenek
Reviewed By

Samantha Korenek
Reviewer Signature

9/9/2026
Date