



|                          |                  |
|--------------------------|------------------|
| Order Number:            | New              |
| Serial # / Local I.D. #: | S2616120003      |
| Calibration Date:        | 9/9/2026         |
| Recommended Due Date:    | 9/9/2027         |
| Certificate Number:      | S261612000346274 |
| Condition Found:         | NEW              |
| Condition Left:          | In Calibration   |

11302 Steeplecrest Dr - Houston, Texas - 77064 - Phone: 832.320.7300 - E-Mail: [www.ofite.com](http://www.ofite.com)

## 294-50 Capillary Suction Timer Calibration Report

| Balance Calibration Check **    |                   |                         |                    |                            |                  |             |
|---------------------------------|-------------------|-------------------------|--------------------|----------------------------|------------------|-------------|
| Target Weight                   | Reading (Seconds) | Temperature (F°)        |                    | Allowed Deviation          | Actual Deviation | Pass / Fail |
| Reading 1                       | 10.1 Sec          | 74.5 F°                 |                    | ± 3.6 Sec                  | 0.8 Sec          | Pass        |
|                                 | Pass              | Pass                    |                    |                            |                  |             |
| Reading 2                       | 10.9 Sec          | 74.5 F°                 |                    | ± 3.6 Sec                  |                  |             |
|                                 | Pass              | Pass                    |                    |                            |                  |             |
| Reading 3                       | 10.8 Sec          | 74.5 F°                 |                    | ± 3.6 Sec                  |                  |             |
|                                 | Pass              | Pass                    |                    |                            |                  |             |
| Calibration Fluid Information * |                   | Calibration Fluid Lot # |                    | Calibration Fluid Exp.Date |                  |             |
| 10000 µS                        |                   | 6061925                 |                    | 6/30/2027                  |                  |             |
| Thermometer I.D.                |                   | 668                     | Certificate Number |                            | C-58767          |             |
| Model Number                    |                   | 90080-09                |                    | Calibration Date           |                  | 1/6/2026    |
| Serial Number                   |                   | 181334600               |                    | Calibration Due Date       |                  | 1/6/2027    |

\* Total dissolved of solids = 6600 Parts Per Million (PPM)

\*\* Three tests of the same temperature should produce a deviation of no more than ± 3.6 seconds.

Emad Hanna  
Calibration Technician

*Emad Hanna*  
Technician Signature

9/9/2026  
Date

Samantha Korenek  
Reviewed By

*Samantha Korenek*  
Reviewer Signature

9/9/2026  
Date