

☐ AS FOUND CHART

☒ AS LEFT CHART




OFI TESTING EQUIPMENT, INC.

11302 Steeplecrest Drive, Houston, Texas 77065 Phone (832) 320-7300 Fax (713) 880-9886 Web www.ofite.com

CALIBRATION CERTIFICATE

Temperatures are calibrated to meet or exceed the requirements stated in API RP 13B-1 and API RP 13B-2

LAB TEMPERATURE	LAB HUMIDITY	PART NUMBER	SERIAL NUMBER	SET POINT	RESULT	ALLOWABLE RANGE	CALIBRATION DATE	DUE DATE	TECHNICIAN
74.00 °F	49.00 %	165-14-3	52617980006	930.0 °F	921.3 °F	860-1000 °F	9/8/2026	9/8/2027	SM



Calibration station has been calibrated using instruments which are traceable to NIST.

CAL STATION SERIAL NUMBER: 6B-11	CAL STATION CALIBRATION DATE: 11/04/2025	CAL STATION DUE DATE: 11/04/2026	Channel 1	
AMBIENT MONITOR MODEL: RH300	AMBIENT MONITOR S/N: 728	AMBIENT MONITOR CAL DATE: 04/30/2026	Plot 1	
UNCERTAINTY: ± 2.8°C / 5°F	CALIBRATION PROCEDURE USED: Retort Calibration Procedures API RP 13B-1	AMBIENT MONITOR DUE DATE: 04/30/2027	Plot 2	
			Plot 3	
			Plot 4	

AS FOUND CONDITION:

CALIBRATION RESULT: Accepted