



11302 Steeplecrest Dr. Houston, TX 77065 Phone 713.880.9885 Fax 713.880.9886 www.ofite.com

PRESSURE TEST CERTIFICATION

- #140-00 – Wall-Mount Filter Press - Basic
- #140-10 – Wall-Mount Filter Press with CO₂ Pressure Assembly
- #140-20 – Bench-Mount Filter Press – Basic
- #140-30 – Bench-Mount Filter Press with CO₂ Pressure Assembly
- #140-31 – Bench-Mount Filter Press with Regulator and Hose
- #140-35 – Bench-Mount Filter Press with Nitrogen Regulator and Cylinder
- #140-36 – Bench-Mount Filter Press
- #140-40 – 4-Unit Lab Model Filter Press
- #140-50 – 6-Unit Lab Model Filter Press

This is to certify that the OFITE product as serialized below has been hydraulically tested to 150 psi +/- 10 psi at 75°F (23.9°C) for 60 seconds. This product is designed to operate at temperatures from 40°F (4.4°C) to 125°F (51.7°C) and a maximum allowable working pressure (MAWP) of 100 psi.

The pressure test for the sub-assemblies was conducted in accordance with OFITE's hydraulic test procedure ENG 105 with an NIST traceable gauge.

Serial Number: _____

Part Number: _____

Cell Serial Number(s): _____

Pressure Assembly SN: _____

Assembled By: _____

Tested By: _____

Date: _____